

Certificate of Analysis

Product Code: RVS001

Product: RVS Powder 500g

LOT / Batch: 147553
Date of Manufacture: 06/04/2017
Expiry Date: 06/04/2020

Physical Characteristics:

Dehydrated Medium: Fine Free Flowing Homogeneous

Reconstituted Medium: Deep Blue

Precipitation: Slight precipitate (Disappears on mixing)

Gel Strength (g/cm²): N/A

pH at 25°C: Specification 5.2 +/- 0.2

Microbiological Performance:

| | Specification | Results |
|-----------------------------------|--------------------------|--------------------------|
| Salmonella typhimurium ATCC 14028 | Recovered from 10-100cfu | Conforms |
| Salmonella enteritidis ATCC 13076 | Recovered from 10-100cfu | Conforms |
| Pseudomonas aeruginosa ATCC 27853 | Satisfactory suppression | Satisfactory suppression |
| Escherichia coli ATCC 25922 | Satisfactory inhibition | Satisfactory inhibition |
| Escherichia coli ATCC 8739 | Satisfactory inhibition | Satisfactory inhibition |
| Enterococcus faecalis ATCC 29212 | Satisfactory inhibition | Satisfactory inhibition |
| Enterococcus faecalis NCTC 775 | Satisfactory inhibition | Satisfactory inhibition |
| Staphylococcus aureus ATCC 6538 | Satisfactory inhibition | Satisfactory inhibition |

Product tested in compliance to ISO 11133:2014

Product is prepared for testing according to documented instructions. Cultures used for challenge testing are from approved culture collections and procedures used are industry accepted methodologies.

The information given above is believed to be correct, however, performance is only warranted when the medium is prepared and used as described in the instructions accompanying the goods either in or on the packaging.

Gillian Bradley, Quality Manager

Date: 13/04/2017