

Certificate of Analysis

Product Code: RELM001
Product: Rapid Enrichment Listeria Media 500g
LOT / Batch: 146214
Date of Manufacture: 09 November 2016
Expiry Date: 01 November 2019

Physical Characteristics:

Dehydrated Medium: Fine free flowing homogenous/Buff
Reconstituted Medium: Dark straw with yellow fluorescence
Precipitation: None
Gel Strength (g/cm²): N/A
pH at 25°C: 7.3 Specification 7.2 +/- 0.1

Microbiological Performance:

	Specification	Results
Listeria monocytogenes ATCC 19111	Recovery from 10-100 cfu	Conforms
Listeria innocua NCTC 11288	Recovery from 10-100 cfu	Conforms
Listeria ivanovii NCTC 11846	Recovery from 10-100 cfu	Conforms
Listeria seeligeri NCTC 11856	Recovery from 10-100 cfu	Conforms
Listeria welshimeri NCTC 11857	Recovery from 10-100 cfu	Conforms
Enterococcus faecalis ATCC 29212	Satisfactory inhibition	Satisfactory inhibition
ELISA test performance	Satisfactory	Satisfactory

Product tested using methodologies in ISO 11133:2014

Product is prepared for testing according to documented instructions. Cultures used for challenge testing are from approved culture collections and procedures used are industry accepted methodologies.

The information given above is believed to be correct, however, performance is only warranted when the medium is prepared and used as described in the instructions accompanying the goods either in or on the packaging.

Gillian Bradley, Quality Manager

Date: 17 November 2016



Cert No 11117
ISO 9001

Solus Scientific Solutions Ltd

9 Mansfield Networkcentre, Millennium Business Park, Concorde Way, Mansfield, Nottinghamshire NG19 7JZ

Tel +44 (0)1623 429701 Fax +44 (0)1623 620977 Email enquiries@solusscientific.com

Web www.solusscientific.com

Registered Address: East Kilbride Laboratory, Torus Building, Rankine Avenue, East Kilbride, Glasgow, Lanarkshire G75 0QF Company Reg. No. SC361616, Registered in Scotland VAT No 982 2304 24