

Results

# **Certificate of Analysis**

Product Code: RVS001

**Product:** RVS Powder 500g

LOT / Batch: 145229

Date of Manufacture: 31 August 2016 Expiry Date: 01 August 2019

### **Physical Characteristics:**

**Dehydrated Medium:** Fine Free Flowing Homogeneous

Reconstituted Medium: Deep Blue

Precipitation: Slight precipitate (Disappears on mixing)

Gel Strength (g/cm²): N/A

**pH at 25°C:** 5.2 Specification 5.2 +/- 0.2

#### **Microbiological Performance:**

Salmonella typhimurium ATCC 14028	Recovered from 10-100cfu	Conforms
Salmonella enteritidis ATCC 13076	Recovered from 10-100cfu	Conforms
Pseudomonas aeruginosa ATCC 27853	Satisfactory suppression	Satisfactory suppression
Escherichia coli ATCC 25922	Satisfactory inhibition	Satisfactory inhibition
Escherichia coli ATCC 8739	Satisfactory inhibition	Satisfactory inhibition
Enterococcus faecalis ATCC 29212	Satisfactory inhibition	Satisfactory inhibition
Enterococcus faecalis NCTC 775	Satisfactory inhibition	Satisfactory inhibition

Specification

Product tested in compliance to ISO 11133:2014

Product is prepared for testing according to documented instructions. Cultures used for challenge testing are from approved culture collections and procedures used are industry accepted methodologies.

The information given above is believed to be correct, however, performance is only warranted when the medium is prepared and used as described in the instructions accompanying the goods either in or on the packaging.

## **Gillian Bradley**, Quality Manager

Date: 02 September 2016

#### **Solus Scientific Solutions Ltd**