

## Certificate of Analysis

<b>Product Code:</b>	<b>LIS-0096S</b>
<b>Product:</b>	Solus <i>Listeria</i> ELISA (1 plate)
<b>LOT/Batch:</b>	<b>EL097</b>
<b>Date of Manufacture/Test:</b>	<b>05.02.2018</b>
<b>Expiry Date:</b>	<b>04.02.2019</b>

This batch has been quality control tested to the following criteria:

Batch homogeneity: CV%	< 10%.
A panel of 9 non- <i>Listeria</i> sp.	<u>Satisfactory</u>
A panel of 10 <i>Listeria</i> sp.	<u>Satisfactory</u>

### Performance of In house Reference Standards

Sample (cfu/ml)	Net Signal OD	OD Specification
D9 <i>L. monocytogenes</i> LMG 16780 10 <sup>6</sup>	0.981	>0.200
D9 <i>L. monocytogenes</i> LMG 16780 10 <sup>7</sup>	>2.50	>1.500
C10 <i>L. monocytogenes</i> ATCC 13932 10 <sup>6</sup>	0.386	>0.200
C10 <i>L. monocytogenes</i> ATCC 13932 10 <sup>7</sup>	1.735	>1.500
D8 <i>L. monocytogenes</i> ATCC 19114 10 <sup>6</sup>	0.765	>0.200
D8 <i>L. monocytogenes</i> ATCC 19114 10 <sup>7</sup>	>2.50	>1.500
D7 <i>L. monocytogenes</i> ATCC 19111 10 <sup>6</sup>	0.308	>0.200
D7 <i>L. monocytogenes</i> ATCC 19111 10 <sup>7</sup>	2.046	>1.500
G3 <i>L. innocua</i> (wild type) 10 <sup>6</sup>	0.202	>0.200
G3 <i>L. innocua</i> (wild type) 10 <sup>7</sup>	1.520	>1.500

#### Solus Scientific Solutions Ltd

9 Mansfield Networkcentre, Millennium Business Park, Concorde Way, Mansfield, Nottinghamshire NG19 7JZ

**Tel** +44 (0)1623 429701 **Fax** +44 (0)1623 620977 **Email** enquiries@solusscientific.com

**Web** www.solusscientific.com

**Registered Address:** East Kilbride Laboratory, Torus Building, Rankine Avenue, East Kilbride, Glasgow, Lanarkshire G75 0QF Company Reg. No. SC361616, Registered in Scotland - VAT No 982 2304 24

Net Optical Densities of samples grown in RELM at 30.0°C for 24 hours and boiled at 85-100°C for 20 minutes.

	1	2	3
A	Negative control 0.058	<i>D5 L. ivanovii</i> >2.500	<i>E9 Bacillus cereus</i> 0.056
B	Positive control >2.500	<i>D9 L. monocytogenes</i> >2.500	<i>A8 E. coli</i> 0.046
C	Substrate blank 0.054	<i>D8 L. monocytogenes</i> >2.500	<i>C1 Staphylococcus aureus</i> 0.091
D	<i>C10 L. monocytogenes</i> >2.500	<i>D7 L. monocytogenes</i> >2.500	<i>E2 E. coli (WT)</i> 0.043
E	<i>G10 L. innocua</i> >2.500	<i>D10 L. seeligeri</i> >2.500	<i>E8 Brochothrix thermosphacta</i> 0.044
F	<i>D2 L. grayii (BHIB)</i> 1.750	<i>E7 Kurthia zopfii</i> 0.045	<i>E6 Enterococcus faecalis</i> 0.050
G	<i>D1 L. monocytogenes</i> >2.500	<i>F1 Streptococcus bovis</i> 0.052	Absorbances read at 450nm
H	<i>D6 L. welshimeri</i> >2.500	<i>F10 Enterococcus faecalis</i> (WT) 0.056	

Production Batch is approved for release as it gives satisfactory QC performance.

Signed: Operations director/Technical director/CEO



Date:

14/3/18

**Solus Scientific Solutions Ltd**

9 Mansfield Networkcentre, Millennium Business Park, Concorde Way, Mansfield, Nottinghamshire NG19 7JZ  
**Tel** +44 (0)1623 429701 **Fax** +44 (0)1623 620977 **Email** enquiries@solusscientific.com  
**Web** www.solusscientific.com

**Registered Address:** East Kilbride Laboratory, Torus Building, Rankine Avenue, East Kilbride, Glasgow, Lanarkshire G75 0QF Company Reg. No. SC361616, Registered in Scotland VAT No 982 2304 24