



**Buffered Peptone Water (BPW) (ISO)      Lot No: 111597**  
**Catalog No. MED017-10KG                      Exp: 28/02/2023**

## PHYSICAL CHARACTERISTICS

<u>SPECIFICATION</u>	<u>EXPECTED RESULTS</u>	<u>ACTUAL RESULTS</u>
<u>Dehydrated:</u> Appearance	Homogenous, free-flowing powder	Homogenous, free-flowing powder
Color	Beige	Beige
<u>Prepared:</u> Clarity	Clear, no precipitate	Clear, no precipitate
Color	Pale yellow	Pale yellow
<u>pH:</u>	7.0 ± 0.2 @ 25°C	7.11 @ 25°C

## CULTURAL RESPONSE- tested in compliance to ISO 11133:2014

The medium was prepared according to label directions and inoculated with the organisms listed below. Cultures were incubated at appropriate atmosphere and temperature and examined for growth at 16-20 hours.

<u>MICROORGANISM</u>	<u>ATCC</u>	<u>APPROX. INOCULUM (CFU)</u>	<u>EXPECTED RESULTS</u>	<u>ACTUAL RESULTS</u>
Pre-enrichment for <i>Salmonella</i> and <i>Enterobacteriaceae</i> detection				
<i>Escherichia coli</i>	25922	10-100	Growth	Meets Expected Result
<i>Escherichia coli</i>	8739	10-100	Growth	Meets Expected Result
<i>Salmonella enteritidis</i>	13076	10-100	Growth	Meets Expected Result
<i>Salmonella typhimurium</i>	14028	10-100	Growth	Meets Expected Result
<i>Cronobacter sakazakii</i>	29544	10-100	Growth	Meets Expected Result
<i>Cronobacter muytjensii</i>	51329	10-100	Growth	Meets Expected Result
Diluent for enumeration of microorganisms and <i>L. monocytogenes</i>				
<i>Escherichia coli</i>	25922	>10 <sup>4</sup>	T <sub>1</sub> plate counts w/in ± 30% of counts for T <sub>0</sub>	Meets Expected Result
<i>Escherichia coli</i>	8739	>10 <sup>4</sup>	T <sub>1</sub> plate counts w/in ± 30% of counts for T <sub>0</sub>	Meets Expected Result
<i>Staphylococcus aureus</i>	25923	>10 <sup>4</sup>	T <sub>1</sub> plate counts w/in ± 30% of counts for T <sub>0</sub>	Meets Expected Result
<i>Listeria monocytogenes</i>	13932	>10 <sup>4</sup>	T <sub>1</sub> plate counts w/in ± 30% of counts for T <sub>0</sub>	Meets Expected Result
<i>Listeria monocytogenes</i>	35125	>10 <sup>4</sup>	T <sub>1</sub> plate counts w/in ± 30% of counts for T <sub>0</sub>	Meets Expected Result



**Buffered Peptone Water (BPW) (ISO)      Lot No: 111597**  
**Catalog No. MED017-10KG                      Exp: 28/02/2023**

**ANIMAL DERIVED MATERIAL COUNTRY OF ORIGIN INFORMATION**

<b><u>ANIMAL DERIVED MATERIAL</u></b>	<b><u>TISSUE SOURCE/ RISK CATEGORY*</u></b>	<b><u>ANIMAL SPECIES</u></b>	<b><u>COUNTRY OF ORIGIN</u></b>
Casein	Milk/ B*	Bovine	<i>Australia, New Zealand</i>
Enzyme	Pancreas/ NA**	Porcine	<i>United States, Canada</i>

\* Risk Categories as defined by the European Pharmacopeia.

\* Not Applicable. "Non-TSE-relevant animal species."

Product was tested in compliance to the regulation noted above. Where a regulation is not referenced, Neogen Corporation standard test methods were observed. Variances in data may occur. External factors such as adverse storage conditions or contamination of product may cause loss in performance. Neogen Corporation certifies that this lot met all quality control specifications for this product.

Date: 07/03/2019

**Gillian Bradley, Quality Manager.**