Certificate of Analysis

Product Code:	**RELM001**
Product:	Rapid Enrichment Listeria Media 500g
LOT / Batch:	**205082**
Date of Manufacture:	**07/11/2018**
Expiry Date:	**06/11/2021**

Physical Characteristics:

Dehydrated Medium:	Fine free flowing homogenous/Buff	
Reconstituted Medium:	Dark straw with yellow fluorescence	
Precipitation:	None	
Gel Strength (g/cm^2):	N/A	
pH at 25°C:	**7.2**	Specification 7.2 +/- 0.1

Microbiological Performance:

	Specification	**Results**
Listeria monocytogenes ATCC 19111	Recovery from 10-100 cfu	Conforms
Listeria innocua NCTC 11288	Recovery from 10-100 cfu	Conforms
Listeria ivanovii NCTC 11846	Recovery from 10-100 cfu	Conforms
Listeria seeligeri NCTC 11856	Recovery from 10-100 cfu	Conforms
Listeria welshimeri NCTC 11857	Recovery from 10-100 cfu	Conforms
Enterococcus faecalis ATCC 29212	Satisfactory inhibition	Satisfactory inhibition
ELISA test performance	Satisfactory	Satisfactory

Product tested using methodologies in ISO 11133:2014

Product is prepared for testing according to documented instructions. Cultures used for challenge testing are from approved culture collections and procedures used are industry accepted methodologies.

The information given above is believed to be correct, however, performance is only warranted when the medium is prepared and used as described in the instructions accompanying the goods either in or on the packaging.

Gillian Bradley, Quality Manager

Date: 16/11/2018

Solus Scientific Solutions Ltd
9 Mansfield Networkcentre, Millennium Business Park, Concorde Way, Mansfield, Nottinghamshire NG19 7JZ
Tel +44 (0)1623 429701 **Fax** +44 (0)1623 620977 **Email** enquiries@solusscientific.com
Web www.solusscientific.com
Registered Address: East Kilbride Laboratory, Torus Building, Rankine Avenue, East Kilbride, Glasgow, Lanarkshire G75 0QF Company Reg. No. SC361616, Registered in Scotland VAT No 982 2304 24