

## Buffered Peptone Water (BPW) (ISO) Lot No: 206262 Catalog No. MED017-10KG

## Exp: Feb 10, 2023

### **PHYSICAL CHARACTERISTICS**

<b>SPECIFICATION</b>		EXPECTED RESULTS	ACTUAL RESULTS
Dehydrated:	Appearance	Homogenous, free-flowing powder	Homogenous, free-flowing powder
	Color	Beige	Beige
Prepared:	Clarity	Clear, no precipitate	Clear, no precipitate
	Color	Pale yellow	Pale yellow
<u>рН</u> :		$7.0 \pm 0.2$ @ $25^{\circ}$ C	$7.0 \pm 0.2$ @ $25^{\circ}$ C

#### CULTURAL RESPONSE- tested in compliance to ISO 11133:2014

The medium was prepared according to label directions and inoculated with the organisms listed below. Cultures were incubated at appropriate atmosphere and temperature and examined for growth at 16-20 hours.

		<u>APPROX.</u> INOCULUM					
<b>MICROORGANISM</b>	ATCC	<u>(CFU)</u>	EXPECTED RESULTS	ACTUAL RESULTS			
Pre-enrichment for Salmonella and Enterobacteriaceae detection							
Escherichia coli	25922	10-100	Growth	Meets Expected Result			
Escherichia coli	8739	10-100	Growth	Meets Expected Result			
Salmonella enteritidis	13076	10-100	Growth	Meets Expected Result			
Salmonella typhimurium	14028	10-100	Growth	Meets Expected Result			
Cronobacter sakazakii	29544	10-100	Growth	Meets Expected Result			
Cronobacter muytjensii	51329	10-100	Growth	Meets Expected Result			
Diluent for enumeration of microorganisms and L. monocytogenes							
Escherichia coli	25922	>104	$T_1 \ plate \ counts \ w/in \pm 30\% \\ of \ counts \ for \ T_0$	Meets Expected Result			
Escherichia coli	8739	>104	$T_1 \ plate \ counts \ w/in \pm 30\% \\ of \ counts \ for \ T_0$	Meets Expected Result			
Staphylococcus aureus	25923	>104	$T_1 \ plate \ counts \ w/in \pm 30\% \\ of \ counts \ for \ T_0$	Meets Expected Result			
Listeria monocytogenes	13932	>104	$T_1 plate \ counts \ w/in \pm 30\% \\ of \ counts \ for \ T_0$	Meets Expected Result			
Listeria monocytogenes	35125	>104	$T_1$ plate counts w/in $\pm$ 30% of counts for $T_0$	Meets Expected Result			



# Buffered Peptone Water (BPW) (ISO)Lot No: 206262Catalog No. MED017-10KGExp: Feb 10, 2023

#### ANIMAL DERIVED MATERIAL COUNTRY OF ORIGIN INFORMATION

<u>ANIMAL DERIVED</u> <u>MATERIAL</u>	<u>TISSUE SOURCE/</u> RISK CATEGORY*	ANIMAL SPECIES	COUNTRY OF ORIGIN
Casein	Milk/ B*	Bovine	Australia, New Zealand
Enzyme	Pancreas/ NA**	Porcine	United States, Canada

\* Risk Categories as defined by the European Pharmacopeia.

\* Not Applicable. "Non-TSE-relevant animal species."

Product was tested in compliance to the regulation noted above. Where a regulation is not referenced, Neogen Corporation standard test methods were observed. Variances in data may occur. External factors such as adverse storage conditions or contamination of product may cause loss in performance. Neogen Corporation certifies that this lot met all quality control specifications for this product.

Date: Monday, February 18, 2019

Gillian Bradley, Quality Manager.