Certificate of Analysis



Lot No.: 301798

Rapid Enrichment Listeria Media 500g

Catalogue No. RELM001 Exp. Date: OCT.09.2023

PHYSICAL CHARACTERISTICS

SPECIFICATION		EXPECTED RESULTS	ACTUAL RESULTS
<u>Dehydrated</u> :	Appearance	Homogenous, free-flowing powder	Homogenous, free-flowing powder
	Colour	Beige	Beige
Prepared:	Appearance	Clear, no precipitate	Clear, no precipitate
	Colour	Dark Yellow with Yellow Fluorescence	Dark Yellow with Yellow Fluorescence
<u>pH</u> :		7.2 ± 0.1 @ 25°C	7.3 @ 25°C

CULTURAL RESPONSE - tested in accordance to ISO 11133:2014

The medium was prepared according to label directions and inoculated with the organisms listed below. Cultures were incubated aerobically at 30 ± 1 °C and examined for growth at 24 hours.

		INOCULUM		
MICROORGANISM	Culture ID	(CFU)	EXPECTED RESULTS	ACTUAL RESULTS
Listeria monocytogenes	ATCC 19111	10-100	Recovery from 10-100 cfu	Meets Expected Result
Listeria innocua	NCTC 11288	10-100	Recovery from 10-100 cfu	Meets Expected Result
Listeria ivanovii	ATCC 19119	10-100	Recovery from 10-100 cfu	Meets Expected Result
Listeria seeligeria	NCTC 11856	10-100	Recovery from 10-100 cfu	Meets Expected Result
Listeria welshimeri	NCTC 18857	10-100	Recovery from 10-100 cfu	Meets Expected Result
Enterococcus faecalis	ATCC 29212	>1000cfu	Satisfactory inhibition	Meets Expected Result
ELISA test performance		N/A	Satisfactory	Meets Expected Result

ANIMAL DERIVED MATERIAL COUNTRY OF ORIGIN INFORMATION

ANIMAL DERIVED MATERIAL	TISSUE SOURCE/ RISK CATEGORY*	ANIMAL SPECIES	COUNTRY OF ORIGIN	CERTIFICATE OF SUITABILITY
Animal Tissue	Lung / B*	Bovine	New Zealand	N/A
Beef Extract	eletal Muscle/ Marrow/ B*	Bovine	New Zealand	N/A
Casein	Milk/ B*	Bovine	New Zealand	N/A
Enzyme	Pancreas/ NA**	Porcine	United States	N/A



Certificate of Analysis



Rapid Enrichment Listeria Media 500g Catalogue No. RELM001 Lot No. 301798

Exp. Date: OCT.09.2023

Product was tested in compliance to the regulation noted above. Where a regulation is not referenced, Solus Scientific Solutions Ltd standard test methods were observed. Variance in data may occur. External factors such as adverse storage conditions or contamination of product may cause loss in performance. Solus Scientific Solutions Ltd certifies that this lot met all quality control specifications for this product.

Gillian Bradley, Quality Manager

Date: Friday, October 23, 2020

