

Certificate of Analysis

Product Code:	LIS1-0480
Product:	Solus ONE <i>Listeria</i> ELISA (5 plate)
LOT/Batch:	NDL006
Date of Manufacture/Test:	26.03.2018
Expiry Date:	25.03.2019

This batch has been quality control tested to the following criteria:

Batch homogeneity: CV%	< 10%.
A panel of 10 non- <i>Listeria spp.</i>	<u>Satisfactory</u>
A panel of 9 <i>Listeria spp</i>	<u>Satisfactory</u>

Performance of In house Reference Standards

Sample	Net Signal OD	OD Specification
<i>L. monocytogenes</i> 4a LMG16783 5x10 ⁴	>2.500	>0.250
<i>L. monocytogenes</i> 4a LMG16783 10 ⁵	>2.500	>0.500
<i>L. monocytogenes</i> 4a LMG16783 10 ⁶	>2.500	>2.000
<i>L. innocua</i> wild type 5x10 ⁴	>2.500	>0.250
<i>L. innocua</i> wild type 10 ⁵	>2.500	>0.500
<i>L. innocua</i> wild type 10 ⁶	>2.500	>2.000
<i>L. welshimeri</i> LMG11389 5x10 ⁴	>2.500	>0.250
<i>L. welshimeri</i> LMG11389 10 ⁵	>2.500	>0.500
<i>L. welshimeri</i> LMG11389 10 ⁶	>2.500	>2.000