

Certificate of Analysis

Product Code: SOL001
Product: SOLO+ 500g
LOT / Batch: 209494
Date of Manufacture: 09/10/2019
Expiry Date: 08/10/2022

Physical Characteristics:

Dehydrated Medium: Fine free flowing homogenous/Buff **Reconstituted Medium:** Dark straw with yellow fluorescence

Precipitation: None Gel Strength (g/cm²): N/A pH at 25°C: 7.2

Specification 7.2 +/- 0.1

Microbiological Performance:

	Specification	Results
Listeria monocytogenes ATCC 19111	Recovery from 10-100 cfu	Conforms
Listeria innocua NCTC 11288	Recovery from 10-100 cfu	Conforms
Listeria ivanovii NCTC 11846	Recovery from 10-100 cfu	Conforms
Listeria seeligeri NCTC 11856	Recovery from 10-100 cfu	Conforms
Listeria welshimeri NCTC 11857	Recovery from 10-100 cfu	Conforms
Enterococcus faecalis ATCC 29212	Satisfactory inhibition	Satisfactory inhibition
ELISA test performance	Satisfactory	Satisfactory

Product tested using methodologies in ISO 11133:2014

Product is prepared for testing according to documented instructions. Cultures used for challenge testing are from approved culture collections and procedures used are industry accepted methodologies.

The information given above is believed to be correct, however, performance is only warranted when the medium is prepared and used as described in the instructions accompanying the goods either in or on the packaging.

Gillian Bradley, Quality Manager

Date: 28/10/2019