

## Certificate of Analysis

**Product Code:** EC-0485  
**Product:** Solus *E. coli* O157 ELISA (1 plate)  
**LOT/Batch:** EE026  
**Date of Manufacture/Test:** 01.07.2019  
**Expiry Date:** 30.06.2020

This batch has been quality control tested to the following criteria:

Batch homogeneity CV%: < 10%.  
 A panel of 8 non- *E. coli* O157 *Enterobacteriaceae* sp: Satisfactory  
 Non toxigenic *E. coli* O157 NCTC 12900: Satisfactory

### Performance of In house Reference Standards

Sample	Net Signal OD	Specification OD
F2 <i>E. coli</i> O157 ~10 <sup>4</sup> cfu/ml	0.771	>0.300
F2 <i>E. coli</i> O157 ~10 <sup>5</sup> cfu/ml	1.367	>0.500
F2 <i>E. coli</i> O157 ~10 <sup>6</sup> cfu/ml	1.445	>1.000

Net Optical Densities read at 450nm of samples grown in mTSB + Novobiocin (20mg/L) for 20 hours at 37°C

	1	2
A	Negative control 0.004	<i>E5 Serratia liquifaciens</i> 0.011
B	Positive control 2.655	<i>B7 Salmonella Typhimurium</i> 0.003
C	Substrate blank 0.046	<i>A8 E. coli</i> 0.007
D	<i>F2 E. coli</i> O157 NCTC 12900 >2.500	<i>E2 E. coli</i> (Wild Type) 0.013
E	<i>A9 Proteus mirabilis</i> 0.003	Absorbances read at 450nm
F	<i>A6 Citrobacter freundii</i> 0.001	
G	<i>A10 Klebsiella pneumoniae</i> 0.005	
H	<i>E4 Shigella sonnei</i> 0.002	

Production Batch is approved for release as it gives satisfactory QC performance.

Signed: Operations director/Technical director/CEO



Date: 8/7/19

**Solus Scientific Solutions Ltd**

9 Mansfield Networkcentre, Millennium Business Park, Concorde Way, Mansfield, Nottinghamshire NG19 7JZ  
 Tel +44 (0)1623 429701 Fax +44 (0)1623 620977 Email enquiries@solusscientific.com  
 Web www.solusscientific.com

**Registered Address:** East Kilbride Laboratory, Torus Building, Rankine Avenue, East Kilbride, Glasgow,  
 Lanarkshire G75 0QF Company Reg. No. SC361616, Registered in Scotland VAT No 982 2304 24