

Certificate of Analysis

Product Code: AGAR-OA020S
Product: Solus Listeria Agar-OA

 LOT/Batch:
 211132

 Date of Manufacture/Test:
 30/01/2020

 Expiry Date:
 23/04/2020

Physical Characteristics	Accredited Method Ref	Specification	Results	
Appearance:	LM not accredited	Straw, opaque agar	Straw, opaque agar	
Fill volume/Weight per pack of 10:	LM3.4	316 -345g	340g	
pH at 25 ⁰ C:	LM3.5	7.2 + 0.2	7.1	
Sterility at 25 & 37 + 1 °C for 5-7 days:	LM3.6	No growth	No growth	

Microbiological Performance:

Organism strain	WDCM	Accredited Method Ref	Specification	Results
Listeria monocytogenes	00109	LM5.13	Recovery >50%, blue, opaque halo	119% Recovery, blue, opaque halo
Listeria monocytogenes	00021	LM5.13	Recovery >50%, blue, opaque halo	110% Recovery, blue, opaque halo
Listeria innocua	00017	LM6.9	Blue, without halo	Blue, without halo
Escherichia coli	00013	LM6.9	Inhibited	Inhibited
Escherichia coli	00012	LM6.9	Inhibited	Inhibited
Enterococcus faecalis	00087	LM6.9	Inhibited	Inhibited
Enterococcus faecalis	00009	LM6.9	Inhibited	Inhibited

Test Methods:

LM3.4 Fill volume weight check

LM3.5 pH test method LM3.6 Sterility test method

LM5.13 Quantitative performance testing of solid medium

LM6.9 Qualitative inoculation of solid medium

Testing relates only to samples representative of the manufactured Lot that have met the specification/acceptance limits on the date of approval. The Uncertainty of Measurement relating to fill volume, pH and microbiological performance has been determined. Cultures used for challenge testing are from approved culture collections and testing is performed in compliance to ISO 11133:2014

The information given above is believed to be correct, however, performance is only warranted when the medium is prepared and used as described in the instructions accompanying the goods either in or on the packaging.

Gillian Bradley, Quality Manager

Date: 07/02/2020

Solus Scientific Solutions Ltd