

Certificate of Analysis



Product Code: MED038
Product: Solus Modified BPW
LOT / Batch: 211972
Date of Manufacture: 03/04/2020
Expiry Date: 02/04/2024

Physical Characteristics:

Dehydrated Medium: Fine Free Flowing Homogenous / Beige
Reconstituted Medium: Pale Yellow
Precipitation: None
pH at 25°C: 7.1 pH specification 7.0 +/- 0.2

Microbiological Performance:

Organism Name	Organism Specification	Result Value
Salmonella typhimurium ATCC 14028	Recovery from 10-100cfu	Recovery from 10-100cfu
Salmonella enteritidis ATCC 13076	Recovery from 10-100cfu	Recovery from 10-100cfu
Escherichia coli ATCC 25922	Recovery from 10-100cfu	Recovery from 10-100cfu
Escherichia coli ATCC 8739	Recovery from 10-100cfu	Recovery from 10-100cfu
Escherichia coli ATCC 25922	+/- 30% colonies/T0	+/- 30% colonies/T0
Escherichia coli ATCC 8739	+/- 30% colonies/T0	+/- 30% colonies/T0
Staphylococcus aureus ATCC 25923	+/- 30% colonies/T0	+/- 30% colonies/T0
Listeria monocytogenes NCTC 10527	+/- 30% colonies/T0	+/- 30% colonies/T0
Listeria monocytogenes NCTC 7973	+/- 30% colonies/T0	+/- 30% colonies/T0

Product tested using methodologies of ISO 11133:2014

Product is prepared for testing according to documented instructions. Cultures used for challenge testing are from approved culture collections and procedures used are industry accepted methodologies.

The information given above is believed to be correct, however, performance is only warranted when the product is prepared and used as described in the current version of the product specification.

Gillian Bradley, Quality Manager

Date: 22/04/2020