

Certificate of Analysis

Product Code:	SAL-04805
Product:	Solus <i>Salmonella</i> ELISA (5 plate)
LOT/Batch:	ES158
Date of Manufacture/Test:	17.02.2020
Expiry Date:	16.02.2021

This batch has been quality control tested to the following criteria:

Batch homogeneity CV%:	< 10%
A panel of 8 non- <i>Salmonella</i> spp:	<u>Satisfactory</u>
A panel of 26 <i>Salmonella</i> spp:	<u>Satisfactory</u>

Performance of In-house Reference Standards

Sample	Net Signal OD	OD Specification
B6 <i>Salmonella</i> Enteritidis 10 ⁶ per ml	1.346	> 1.000
B6 <i>Salmonella</i> Enteritidis 10 ⁷ per ml	>2.500	>2.000
B7 <i>Salmonella</i> Typhimurium 10 ⁶ per ml	0.785	>0.750
B7 <i>Salmonella</i> Typhimurium 10 ⁷ per ml	>2.500	>2.000
B2 <i>Salmonella</i> Nottingham 10 ⁶ per ml	0.348	>0.200
B2 <i>Salmonella</i> Nottingham 10 ⁷ per ml	1.436	>0.750

Solus Scientific Solutions Ltd

9 Mansfield Networkcentre, Millennium Business Park, Concorde Way, Mansfield, Nottinghamshire NG19 7JZ
Tel +44 (0)1623 429701 **Fax** +44 (0)1623 620977 **Email** enquiries@solusscientific.com
Web www.solusscientific.com

Registered Address: East Kilbride Laboratory, Torus Building, Rankine Avenue, East Kilbride, Glasgow, Lanarkshire G75 0QF Company Reg. No. SC361616, Registered in Scotland VAT No 982 2304 24

Net Optical Densities of samples grown in RVS at 41.5°C for 24 hours and boiled at 85-100°C for 20 minutes

	1	2	3	4	5
A	Negative control 0.013	<i>Klebsiella pneumonia</i> 0.009	<i>Salmonella</i> Nima >2.500	<i>Salmonella</i> Ealing 0.899	<i>Salmonella</i> Typhimurium (non- motile wild type) >2.500
B	Positive control >2.500	<i>Shigella sonnei</i> 0.009	<i>Salmonella</i> Essen 2.023	<i>Salmonella</i> Kirkee 1.792	<i>Salmonella</i> Hofit 1.019
C	Substrate blank 0.055	<i>Serratia liquifaciens</i> 0.011	<i>Salmonella</i> Newport (wild type) 1.985	<i>Salmonella</i> Abaetatuba >2.500	<i>Salmonella</i> Menston 1.893
D	<i>Staph. aureus</i> 0.012	<i>Salmonella</i> Heves >2.500	<i>Salmonella</i> Dublin 2.136	<i>Salmonella</i> Bracknell 2.279	<i>Salmonella</i> Minnesota 1.335
E	<i>Citrobacter freundii</i> 0.013	<i>Salmonella</i> Seftenberg >2.500	<i>Salmonella</i> Anatum >2.500	<i>Salmonella</i> Gallinarum (wild type) 2.047	<i>Salmonella</i> Kentucky 2.058
F	<i>Serratia marcescens</i> 0.009	<i>Salmonella</i> Poona >2.500	<i>Salmonella</i> Aschesleben >2.500	<i>Salmonella</i> Tranaroa 2.258	Absorbances read at 450nm
G	<i>E. coli</i> 0.012	<i>Salmonella</i> Hadar >2.500	<i>Salmonella arizonae</i> NCTC 7349 0.255	<i>Salmonella</i> Virchow 1.763	
H	<i>Proteus mirabilis</i> 0.018	<i>Salmonella</i> Dahlem >2.500	<i>Salmonella</i> Nitra >2.500	<i>Salmonella</i> Dugbe >2.500	

Production Batch is approved for release as it gives satisfactory QC performance.

Signed: Operations director/~~Technical director~~/CEO



Date:

26/2/2020

Solus Scientific Solutions Ltd

9 Mansfield Networkcentre, Millennium Business Park, Concorde Way, Mansfield, Nottinghamshire NG19 7JZ

Tel +44 (0)1623 429701 Fax +44 (0)1623 620977 Email enquiries@solusscientific.com

Web www.solusscientific.com

Registered Address: East Kilbride Laboratory, Torus Building, Rankine Avenue, East Kilbride, Glasgow, Lanarkshire G75 0QF Company Reg. No. SC361616, Registered in Scotland VAT No 982 2304 24