

Certificate of Analysis



Product Code: SALSUPP-22.5
Product: Solus One Salmonella Supplement
LOT / Batch: UG05S
Date of Manufacture: 17/08/2020
Expiry Date: 17/08/2022

Satisfactory growth observed from the test organism in Buffered Peptone Water incubated for 24 h at 37°C.

The supplement when incorporated into Mueller Hinton agar gave the expected growth/inhibition profile with 36 bacteria when incubated at 37°C for 24h.

Sterility of the supplement was confirmed as no growth was observed when added to blood agar incubated for 3 days at 37°C and 20°C.

<u>Performance:</u>	<u>Specification</u>	<u>Results</u>
<i>Salmonella typhimurium</i> NCTC 12023	Growth	Growth
<i>Staphylococcus aureus</i> ATCC 25923	No Growth	No Growth

The performance of this product is based on use of the supplement as detailed in the accompanying instructions for use.

Gillian Bradley, Quality Manager

Date: 17/08/2020