

Certificate of Analysis



Product Code: LIS-0480S
Product: Solus *Listeria* ELISA (5 plate)
LOT / Batch: EL160
Date of Manufacture: 19th July 2021
Expiry Date: 18th October 2022

This batch has been quality control tested to the following criteria:

Batch homogeneity: CV% < 10%.
A panel of 9 non-*Listeria* spp: Satisfactory
A panel of 11 *Listeria* spp: Satisfactory

Performance of In house Reference Standards

Sample (cfu/ml)	Net Signal OD	OD Specification
C10 <i>L. monocytogenes</i> ATCC 13932 10 ⁶	1.072	>0.400
C10 <i>L. monocytogenes</i> ATCC 13932 10 ⁷	>2.500	>2.000
D5 <i>L. ivanovii</i> ATCC 19119 10 ⁶	0.428	>0.250
D5 <i>L. ivanovii</i> ATCC 19119 10 ⁷	2.104	>2.000
D6 <i>L. welshimeri</i> ATCC 35897 10 ⁶	>2.500	>1.500
D6 <i>L. welshimeri</i> ATCC 35897 10 ⁷	>2.500	>2.000
D7 <i>L. monocytogenes</i> ATCC 19111 10 ⁶	0.808	>0.400
D7 <i>L. monocytogenes</i> ATCC 19111 10 ⁷	>2.500	>2.000
G3 <i>L. innocua</i> (wild type) 10 ⁶	0.661	>0.250
G3 <i>L. innocua</i> (wild type) 10 ⁷	>2.500	>2.000

Certificate of Analysis



Net Optical Densities of samples grown in RELM at 30.0°C for 24 hours and boiled at 85-100°C for 20 minutes.

	1	2	3
A	Negative control 0.003	<i>D7 L. monocytogenes</i> >2.500	<i>E8 Brochothrix thermosphacta</i> 0.023
B	Positive control >2.500	<i>D8 L. monocytogenes</i> >2.500	<i>E9 Bacillus cereus</i> 0.023
C	Substrate blank 0.056	<i>D9 L. monocytogenes</i> >2.500	<i>E6 Enterococcus faecalis</i> 0.025
D	<i>G10 L. innocua</i> >2.500	<i>C10 L. monocytogenes</i> >2.500	<i>AT5 Bacillus brevis (Wild type)</i> 0.026
E	<i>D5 L. ivanovii</i> >2.500	<i>D10 L. seeligeri</i> 0.505	<i>C1 Staphylococcus aureus</i> 0.024
F	<i>D6 L. welshimeri</i> >2.500	<i>G3 L. innocua</i> >2.500	<i>AT6 Bacillus cereus (Wild type)</i> 0.023
G	<i>D2 L. grayii</i> >2.500	<i>F1 Streptococcus bovis</i> 0.024	<i>F10 Enterococcus faecalis (Wild type)</i> 0.020
H	<i>D1 L. monocytogenes</i> >2.500	<i>E7 Kurthia zopfii</i> 0.023	Absorbances read at 450nm

Production Batch is approved for release as it gives satisfactory QC performance.

Signed: Technical director/Operations Manager/Quality Manager

Date: 27/7/21.