

Certificate of Analysis



Product Code: EC-048S
Product: Solus *E. coli* O157 ELISA (1 plate)
LOT / Batch: EE034
Date of Manufacture: 10.05.2021
Expiry Date: 09.11.2022

This batch has been quality control tested to the following criteria:

Batch homogeneity CV%: < 10%.
A panel of 8 non - *E. coli* O157 *Enterobacteriaceae* spp: Satisfactory
Non-toxicogenic *E. coli* O157 NCTC 12900: Satisfactory

Performance of In house Reference Standards

Sample	Net Signal OD	Specification OD
F2 <i>E. coli</i> O157 ~10 ⁴ cfu/ml	0.417	>0.300
F2 <i>E. coli</i> O157 ~10 ⁵ cfu/ml	0.976	>0.500
F2 <i>E. coli</i> O157 ~10 ⁶ cfu/ml	1.245	>1.000

Net Optical Densities read at 450nm of samples grown in mTSB + Novobiocin (20mg/L) for 20 hours at 41.5°C

	1	2
A	Negative control 0.008	<i>E5 Serratia liquifaciens</i> 0.011
B	Positive control 1.800	<i>B7 Salmonella Typhimurium</i> 0.006
C	Substrate blank 0.040	<i>A8 E. coli</i> 0.002
D	<i>F2 E. coli</i> O157 NCTC 12900 >2.500	<i>E2 E. coli</i> (Wild Type) 0.015
E	<i>A9 Proteus mirabilis</i> 0.044	Absorbances read at 450nm
F	<i>I3 Citrobacter freundii</i> 0.025	
G	<i>A10 Klebsiella pneumoniae</i> 0.012	
H	<i>E4 Shigella sonnei</i> 0.041	

Certificate of Analysis



Production Batch is approved for release as it gives satisfactory QC performance.

Signed: Technical director/~~Operations Manager~~/Quality Manager

DS.

Date: 17/5/21.