### **Certificate of Analysis**



# FRASER BROTH Catalogue No. FB001

Lot No: 304089

Exp. Date: MAR.08.2023

### PHYSICAL CHARACTERISTICS

<b>SPECIFICATION</b>		EXPECTED RESULTS	ACTUAL RESULTS
<u>Dehydrated</u> :	Appearance	Homogenous, free-flowing powder	Homogenous, free-flowing powder
	Color	Beige	Beige
Prepared:	Clarity	Clear	Clear
	Color	Yellow	Yellow
	Precipitate	None to Slight	None to Slight
<u>pH</u> :		$7.2 \pm 0.2$ @ 25°C	7.1 @ 25°C

CULTURAL RESPONSE – tested in compliance to ISO 11133:2014+A1:2018

#### (Fertility and Selectivity Testing)

The medium was prepared according to label directions and 10mL volumes inoculated with the organisms listed below. Cultures were incubated aerobically at  $37 \pm 1^{\circ}$ C and examined for growth at 22-26 hours. Following incubation,  $10\mu$ L of *E. faecalis* and

E. coli cultures were subcultured onto TSA at  $37 \pm 1$  °C and examined for growth at 22-26 hours

		ADDDOV	EXPECTED RESULTS			ACTUAL RESULTS		
MICROORGANISM	<u>ATCC</u>	APPROX. INOCULUM (CFU)	Growth in FB	Reaction in FB	Recovery on TSA	Growth in FB	Reaction in FB	Recovery on TSA
Enterococcus faecalis	19433	> 10 <sup>3</sup>	Partial to Complete Inhibition		< 100 cfu	Mo	eets Expected	Result
Enterococcus faecalis	29212	> 10 <sup>3</sup>	Partial to Complete Inhibition		< 100 cfu	Mo	eets Expected	Result
Escherichia coli	25922	> 10 <sup>3</sup>	Complete Inhibition		< 10 cfu	Me	eets Expected	Result
Escherichia coli	8739	> 10 <sup>3</sup>	Complete Inhibition		< 10 cfu	Me	eets Expected	Result
Listeria innocua	33090	$> 10^{3}$	Growth	Blackening	NA	Me	eets Expected	Result
Listeria monocytogenes	35152	10-100	Growth	Blackening	NA	Me	eets Expected	Result
Listeria monocytogenes	13932	10-100	Growth	Blackening	NA	Me	eets Expected	Result
Staphylococcus aureus	25923	> 10 <sup>4</sup>	Partial to complete inibition	-	NA	M	eets Expected	Result



## **Certificate of Analysis**



# FRASER BROTH Catalogue No. FB001

(Productivity Testing)

The medium was prepared according to label directions and 10mL volumes inoculated as a mixed culture using the organisms listed below. Cultures were incubated aerobically at  $37 \pm 1^{\circ}$ C, examined for growth at 22-26 hours. and sub cultured onto Listeria Chromogenic Agar\* at  $37 \pm 1^{\circ}$ C and plates examined for growth at 22-50 hours

Lot No: 304089

Exp. Date: MAR.08.2023

	<u>ATCC</u>	APPROX. INOCULUM (CFU)	<b>EXPECTE</b>	D RESULTS	ACTUAL RESULTS	
Mixed Culture Testing			Recovery on Listeria Agar*	Reaction on Listeria Agar*	Recovery on Listeria Agar*	Reaction on Listeria Agar*
Enterococcus faecalis +	29212	$> 10^{3}$	Inhibited	_	Meets Exp	ected Result
Escherichia coli +	25922	$>10^{3}$	Inhibited	_	Meets Expected Result	
Listeria monocytogenes	35152	10-100	>10 cfu	Blue colonies with opaque halo	Meets Expected Result	
Enterococcus faecalis +	19433	$> 10^{3}$	Inhibited	_	Meets Exp	ected Result
Escherichia coli +	8739	$>10^{3}$	Inhibited	_	Meets Expe	ected Result
Listeria monocytogenes	13932	10-100	>10 cfu	Blue colonies with opaque halo	Meets Expo	ected Result

<sup>\*</sup>according to Ottaviani and Agosti

### ANIMAL DERIVED MATERIAL COUNTRY OF ORIGIN INFORMATION

ANIMAL DERIVED MATERIAL Animal Tissue	TISSUE SOURCE/ RISK CATEGORY* Lung/ B*	ANIMAL SPECIES Bovine	COUNTRY OF ORIGIN New Zealand	CERTIFICATE OF SUITABILITY N/A
Casein	Milk/ B*	Bovine	New Zealand	N/A
Enzyme	Pancreas/ NA**	Porcine	United States	N/A
Beef Extract	Skeletal Muscle/Bone /B*	Bovine	New Zealand	R1-CEP 2001-299

 $<sup>\</sup>boldsymbol{*}$  Risk Categories as defined by the European Pharmacopeia.

Product was tested in compliance to the regulation noted above. Variances in data may occur. External factors such as adverse storage conditions or contamination of product may cause loss in performance. Solus Scientific Solutions Ltd certifies that this lot met all quality control specifications for this product.

### Gillian Bradley, Quality Manager

Date: Thursday, March 18, 2021



<sup>\*\*</sup> Not Applicable. "Non-TSE-relevant animal species."