

Certificate of Analysis



Solus Listeria Agar-OA
Catalogue No. AGAR-0A020S

Lot No.: 311009
D.O.M.: NOV.22.2021
Exp. Date: FEB.28.2022

PHYSICAL CHARACTERISTICS

<u>SPECIFICATION</u>	<u>ACCREDITED METHOD</u>	<u>EXPECTED RESULTS</u>	<u>ACTUAL RESULTS</u>
<u>Appearance:</u>	LM not accredited	Straw, opaque agar	Straw, opaque agar
<u>pH at 20-25°C:</u>	LM3.5	7.2 ± 0.2	7.1
<u>Fill volume/weight per pack of 10:</u>	LM3.4	316-345g	331g
<u>Sterility at 25 & 37°C ± 1°C for 5-7 days:</u>	LM3.6	No Growth	No Growth

CULTURAL RESPONSE: tested in compliance to ISO 11133:2014/A2:2020

<u>MICROORGANISM</u>	<u>CULTURE ID</u>	<u>ACCREDITED METHOD</u>	<u>EXPECTED RESULTS</u>	<u>ACTUAL RESULTS</u>
<i>Listeria monocytogenes</i>	WDCM 00109	LM5.13	Recovery >50%, blue, opaque halo	124%, blue, opaque halo
<i>Listeria monocytogenes</i>	WDCM 00021	LM5.13	Recovery >50%, blue, opaque halo	102%, blue, opaque halo
<i>Listeria innocua</i>	WDCM 00017	LM6.9	Blue, without halo	Blue, without halo
<i>Escherichia coli</i>	WDCM 00013	LM6.9	Total inhibition	Total inhibition
<i>Escherichia coli</i>	WDCM 00012	LM6.9	Total inhibition	Total inhibition
<i>Enterococcus faecalis</i>	WDCM 00087	LM6.9	Total inhibition	Total inhibition
<i>Enterococcus faecalis</i>	WDCM 00009	LM6.9	Total inhibition	Total inhibition

Test Methods:

LM3.4 Fill volume weight check
LM3.5 pH test method
LM3.6 Sterility test method
LM5.13 Quantitative performance testing of solid medium
LM6.9 Qualitative inoculation of solid medium

Testing relates only to samples representative of the manufactured lot that have met the specification/acceptance limits on the date of approval. The sampling method is not an accredited method.

The Uncertainty of Measurement relating to fill volume, pH and microbiological performance has been determined.

Cultures used for challenge testing are from approved culture collections and testing is performed as per the requirements for ISO 11133:2014/A2:2020

The information given above is believed to be correct, however, performance is only warranted when the medium is used according to manufacturer's recommendations.

Gillian Bradley, Quality Manager

Date: Thursday, November 29, 2021